



# Calibration Report

Test # 202812



**United States - Main Office**

3839 County Road 116

Medina, MN 55340 USA

www.intercompcompany.com

TEL. (763) 476-2531 FAX (763) 476-2613

**Intercomp Europe**

Manor Farm, Manor Road

Shurlock Row

Berkshire RG10 0PY UK

TEL. 44-118-932-0578 FAX 44-118-932-1034

<b>Tester</b>	CB			<b>Model #</b>	SW650 RFX BLUETOOTH		
<b>Date</b>	03/Jan/2024			<b>Scale/Cell ID #</b>			
<b>Indicator</b>				<b>Capacity</b>	1500 lb		
<b>Serial#</b>	0323MS20005			<b>Graduation (d)</b>	1		
<b>Part #</b>	170319			<b>Re-Cal Date</b>	03/Jan/2025		
<b>Temperature</b>	16.2 °C	<b>Humidity</b>	61 %	<b>Ack #</b>		<b>PO #</b>	

Accuracy: ±0.1 % of applied load or ±1 lb - whichever is greater

**As Left: In Tolerance**


Weight(Lbs.)	Left Front(Lbs.)	Left Rear(Lbs.)	Right Front(Lbs.)	Right Rear(Lbs.)
0	0	0	0	0
200	200	200	200	200
500	500	500	500	500
1000	1000	1000	1000	1000
0	0	0	0	0


Intercomp Company does hereby certify the above listed instrument meets or exceed all published Specifications and has been calibrated using standards whose accuracies are directly traceable to the U.S. National Institute of Standards and Technology.

This document is not to be reproduced except in full without the written permission of Intercomp

**Calibration Procedure:**

Model #	Test/Trace #	Re-Cal Date
CS10-5K-B308	Tovey 54239-1	27/Feb/2024

Tested By:   
 Title: Director \_\_\_\_\_

Approve By:   
 Title: Director \_\_\_\_\_